

# Certificate of Calibration

Certificate Number: 766325



**JJ Calibrations, Inc.**

7724 SE Aspen Summit Drive  
Portland, OR 97266-9217  
Phone 503.786.3005  
FAX 503.786.2994

**VTG, LLC (IND Analytical)**  
520 SE Columbia River Dr  
#223  
Vancouver, WA 98661

PO: 362426

Order Date: 03/04/2022

Authorized By: N/A

Calibrated on: 03/06/2022

\*Recommended Due: 03/06/2023

Environment: 20 °C 36 % RH

As Received: Within Tolerance

As Returned: Within Tolerance

Action Taken: Calibrated

Technician: 40

Property #: N/A  
User: N/A  
Department: N/A  
Make: Prostat  
Model: PRS - 801  
Serial #: 0815A091  
Description: Resistance Meter  
Procedure: 402446  
Accuracy: Refer Spec.

Remarks: \* Many factors may cause the unit to drift out of calibration before the recommended due date. Any reported error is the absolute value between the reference and the unit.

Returned with case, probes and manual.

## Standards Used

<u>Std ID</u>	<u>Manufacturer</u>	<u>Model</u>	<u>Nomenclature</u>	<u>Due Date</u>	<u>Trace ID</u>
720A	IET	HARS-X-11-0.001	High Accuracy Resistance Substi	05/06/2022	742457
711A	JJ Calibrations	HV Resistance Box	High Voltage Resistance Box	05/31/2022	747897
310A	Victoreen	RX-1M-3	Resistor Set, High Meg	07/19/2022	747916

This instrument has been calibrated in accordance with the JJ Calibrations Quality Assurance Manual and is traceable to either the SI or to National Institute of Standards and Technology (NIST). The quality system and this certificate are in compliance with ANSI/NCSL Z540-1-1994, ISO/IEC 17025-2017, ISO 10012-1, the ISO 9000 family and QS 9000. The expanded uncertainties of measurements for this calibration are based upon 95% (2 sigma) confidence limits. Unless stated in the comments, certificates reflect the "Simple Acceptance Rule" as specified by JCGM 106:2012. Unless otherwise stated, a test accuracy ratio (TAR) of 4:1, if achievable, is maintained. The results reported herein apply only to the calibration of the item described above. This report may not be reproduced, except in full, without written approval of JJ Calibrations.

  
\_\_\_\_\_  
Reviewer

Issued 03/08/2022

Rev # 15

  
\_\_\_\_\_  
Inspector