



Sensidyne Certificate of Performance Gilian LFS-113 Sampling Pumps

This document certifies that the product below performs in accordance with factory specifications. Sensidyne's volumetric test equipment is traceable to NIST. Sensidyne, LP is an ISO 9001:2015 registered company.

LFS-113DC Assembly, Clock, P/N 800098-R
Serial Number 20200902005

Month of Manufacture: September 2020

Set Flow cc/min	Set BP Inches H2O	Acceptable Minimum cc/min	Acceptable Maximum cc/min	Pass = √ Fail = X
200	0 25	190 190	210 210	√ √
100	0 25	95 95	105 105	√ √
50	0 25	47.5 47.5	52.5 52.5	√ √

Sensidyne
507

Technician Stamp _____