



SKC Certificate of Compliance Universal XR Sampler

This is to certify that the item listed below is in accordance with factory specifications. SKC test equipment is calibrated in accordance with ISO/IEC 17025 utilizing NIST and/or UKAS traceability standards.

Model Number 224-PCXR8 Serial Number A127U54

Flow ml/min	Settings		Acceptance Criteria		✓
	BP inches of water		Minimum ml/min	Maximum ml/min	
5000	0		5000	5000	✓
	10		4750	5250	✓
4000	0		4000	4000	✓
	20		3800	4200	✓
3000	0		3000	3000	✓
	35		2850	3150	✓
2500	0		2500	2500	✓
	40		2375	2625	✓
2000	0		2000	2000	✓
	20		1900	2100	✓
	40		1900	2100	✓
1000	0		1000	1000	✓
	20		950	1050	✓
	40		950	1050	✓

Check Points	✓
Set Regulator 20"	✓
Keyboard Check	✓
Flow Fault	✓

Technician # 50241

Scott Marshall
Scott Marshall

Quality Assurance Manager