

**Certificate No. 0190-H00033-06-2022**

**Equipment**

Type: Portable/Fixed/Bench-Top  
Manufacturer: Meeco  
Model: Tracer 2  
Serial No. 20211D-44  
Internal Catalog No.  
Measuring Components: Moisture  
Measuring Range: 0-2 ppm  
Accuracy: better than 2% of measuring range

**Customer: QPLUS**

Executed by: Hagai Liviatan  
Date: 19.06.2022  
Next Calibration: 06-2023  
Signature: \_\_\_\_\_

1. Visual OK: Passed / Failed
2. Comments: \_\_\_\_\_

**Calibration Gas**

	Concentration	Manufacturer	Due Date	Cyl. Number
Zero	99.99995% N2	CalGaz	05.2027	955-512314
	Incl. Purifier			

**Results**

Component				
Zero Before	48 ppb			
Span Before				
Zero After	22 ppb			
Span After				
Zero Error	.			
Span Error				

**Analog Output**


Required/Not Required	
Value 1	mA
Value 2	mA

**HMI**

Required/Not Required	
Equipment	HMI

Comments: Visual and sound alarm were tested. Operated well.

**Final Result: Pass / Fail**





Hagai Liviatan | CEO

Tel: +972-542280140

Email: hagail@liviatan.com

Technologies Ltd.

Web: [www.liviatantechnologies.com](http://www.liviatantechnologies.com)



**Certificate No. 0191-H00034-06-2022**

**Equipment**

Type: Portable/Fixed/Bench-Top  
Manufacturer: AII  
Model. PI2-MS-500  
Serial No. 7172641  
Measuring Components: Oxygen  
Measuring Range: 0-500 ppb  
Accuracy: better than 2% of measuring range

**Customer: QPLUS**

Executed by: Hagai Liviatan  
Date: 19.06.2022  
Next Calibration: 12-2022  
Signature: \_\_\_\_\_

1. Visual OK: Passed / Failed

2. Comments: \_\_\_\_\_  
\_\_\_\_\_

**Calibration Gas**

	Concentration	Manufacturer	Due Date	Cyl. Number
Zero	99.99995% N2	CalGaz	05.2027	955-512314
Span	8 ppm O2 in 99.999% N2	Linde	04.2025	A431-F4

**Results**

Component				
Zero Before	90 ppb			
Span Before	7.988 ppm			
Zero After	34 ppb			
Span After	8 ppm			
Zero Error	.			
Span Error				

**Analog Output**

Required/Not Required

Value 1	mA
Value 2	mA

**HMI**

Required/Not Required

Equipment	HMI

Comments: Visual and sound alarm were tested. Operated well.

**Final Result: Pass / Fail**

